

CLARIFICATIONS no. 1

Answers to the questions raised by the bidders:

Request for Quotation
Supply of Medical Science Lab Equipment for Secondary School
Ref: IOM/RFQ/2024/025

Nr.		
1.	Please kindly confirm that all dimensions where being requested can be subject to 15% variation	We would prefer compliance to required specifications.
2.	Item 4 Hospital mechanical bed, Height: 48 cm. Please clarify If the required height is the height of the section incline?	It is not the height of the incline, but it is the height from the floor to the horizontal part of the bed.
3.	Item 10 is described as Sterilizer E-220 22L – however, in the further description of the equipment, it is required “Internal Volume 16 Lt., Internal Dimensions Height x Width x Depth 300 x 250 x 220 mm”. Please confirm a technical mistake is being made and a Sterilizer as of 16 Lt. is being required	A technical error has occurred, the model is Sterilizer M-160 “or equivalent”, and the 16l model is required
4.	Item Medical record cabinet: Please confirm if double metal doors will be acceptable? Please confirm that Medical record cabinet with lightgray colour will be acceptable?	The double metal doors are acceptable only for article Nr 15. The light gray color is acceptable
5.	Item 38 Stainless steel medical waste bin 2L: Please clarify if a technical mistake is being made regarding the volume of the bin? Please note that minimum volume of Stainless steel medical waste bins of inox and with a pedal is 3l. Another option is a plastic bin, without pedal for needles, sharps and cutting materials, medical waste with possibility to be burnt into an incinerator. Please clarify if a 3l Stainless steel medical waste bin with	Acceptable is a metal pedal bin with a minimum volume of 2L (can be 3L or more)

	pedal or a 2l plastic medical waste bin without pedal will be acceptable.	
6.	Items 27 to 35 – could you please clarify if you will require 2 pcs of each item only, or 2 boxes (25 pcs each).	It is 2 pieces of each

EoD.